



ASO

Fast Test Kit

(Immunofluorescence Assay)

IF1076 for Getein1100
 IF5076 for Getein1160
 IF3076 for Getein1180
 IF4076 for Getein1200
 IF2076 for Getein1600



User Manual

INTENDED USE

ASO Fast Test Kit (Immunofluorescence Assay) is intended for in vitro quantitative determination of anti streptolysin "O" antibody (ASO) in serum, plasma or whole blood samples. ASO is a very valuable marker for the diagnosis of group A streptococcal infection, its titer can reflect the severity of infection. Some diseases will lead to a significant increase in ASO titer, such as acute rheumatic fever.

SUMMARY

Streptolysin "O", a protein with hemolytic activity, is one of the important metabolites of group A streptococcus. It can dissolve the red blood cells of human and some animals. Streptolysin "O" has strong antigenicity. When human is infected with group A hemolytic streptococcus, B lymphocytes will secrete corresponding antibodies under the stimulation of streptolysin "O", that is, anti streptolysin "O" (ASO). This antibody can neutralize streptolysin "O" and disable its hemolytic ability.

ASO began to rise one week after group A Streptococcus infection, peaked in 4-5 weeks and lasted for several months. As the infection subsides, ASO decreased and returned to normal value within 6 months. The disease condition can be judged by repeated measurements over time. The increase of ASO over time indicates the early stage of infection, decrease indicates the subside stage of infection. The lack of decline in ASO titer suggested the possibility of recurrent or chronic infection.

PRINCIPLE

The test kit adopts a double-antigen sandwich method to quantitatively detect the concentration of ASO in human serum, plasma and whole blood samples. After the sample has been applied to the test card, the fluorescence latexlabelled ASO antigen binds with ASO in sample and forms a marked antigen-antibody complex. The

complex moves to the detection area by capillary action, then it is captured by ASO antigen coated on the detection area of nitrocellulose membrane, forming a double-antigen complex. The complex generates a fluorescent signal and the intensity increases in proportion to the amount of ASO in sample. Then insert test card into Getein1100/Getein1160/Getein1180 Immunofluorescence Quantitative Analyzer/ automatically inserted by Getein1200/Getein1600 Immunofluorescence Quantitative Analyzer (hereinafter referred to as Getein1100, Getein1160, Getein1180, Getein1200 and Getein1600), the concentration of ASO in sample will be measured and displayed on the screen. The value will be stored in Getein1100/Getein1160/Getein1180/Getein1200/Getein1600 and available for downloading. The result can be easily transmitted to the laboratory or hospital information system.

CONTENTS

1. A kit for Getein1100/Getein1160/Getein1180 contains:

- Package specifications: 25 tests/kit, 10 tests/kit
- 1) Getein ASO test card in a sealed pouch with desiccant
 - 2) Capillary pipet
 - 3) Sample diluent
 - 4) User manual: 1 piece/kit
 - 5) SD card: 1 piece/kit

2. A kit for Getein1200/Getein1600 contains:

- Package specifications: 2x24 tests/kit, 2x48 tests/kit
- 1) Sealed cartridge with 24/48 Getein ASO test cards
 - 2) User manual: 1 piece/kit
- Materials required for Getein1200/Getein1600:

- 1) Sample diluent: 1 bottle/kit
- 2) Box with pipette tips: 96 tips/kit
- 3) Mixing plate: 1 piece/kit

3. A test card consists of:

A plastic shell and a reagent strip which is composed of a sample pad, nitrocellulose membrane (the test line is coated with ASO antigen and the control line is coated with polyclonal mouse anti streptolysin "O" antibody), fluorescence latex-labelled ASO antigen, absorbent paper and liner.

4. Sample diluent composition:

Phosphate buffered saline, proteins, detergent, preservative, stabilizer.

Note: Do not mix or interchange different batches of kits.

APPLICABLE DEVICE

Getein1100 Immunofluorescence Quantitative Analyzer
 Getein1160 Immunofluorescence Quantitative Analyzer
 Getein1180 Immunofluorescence Quantitative Analyzer
 Getein1200 Immunofluorescence Quantitative Analyzer

Getein1600 Immunofluorescence Quantitative Analyzer

STORAGE AND STABILITY

Store the test card at 4–30°C with a valid period of 24 months.

Use the test card for Getein1100/Getein1160/Getein1180 within one hour once the foil pouch is opened. For test card of Getein1200/Getein1600: if the cartridge is opened, it could be stable within 24 hours once exposed to air. If the test cards can't be used up at a time, please put the cartridge back to the foil pouch and reseal along the entire edge of zip-seal. The remaining test cards should be used up within 7 days.

PRECAUTIONS

1. For in vitro diagnostic use only.
2. For professional use only.
3. Do not use the kit beyond the expiration date.
4. Do not use the test card if the foil pouch or the cartridge is damaged.
5. Do not open pouches or the cartridge until ready to perform the test.
6. Do not reuse the test card.
7. Do not reuse the pipet.
8. Handle all specimens as potentially infectious. Proper handling and disposal methods should be followed in accordance with local regulations.
9. Carefully read and follow user manual to ensure proper test performance.

SPECIMEN COLLECTION AND PREPARATION

1. **Serum, plasma or whole blood** samples can be used for the test. Other body fluids and samples may not give accurate results. Samples should be free of hemolysis.
2. Venous blood should be collected under aseptic conditions; serum or plasma is preferred for testing.
3. Heparin, sodium citrate or EDTA can be used as the anticoagulant for plasma and whole blood samples.
4. The test should be performed at room temperature within 4 hours after sample collection.
5. If testing is delayed, serum and plasma samples may be stored up to 7 days at 2–8°C and 3 months at -20°C before testing. Whole blood samples should not be frozen and can be stored at 2–8°C for 3 days. Do not heat inactivated samples or use hemolyzed blood samples.
6. Refrigerated or frozen sample should be reached to room temperature before testing. Frozen samples must be completely thawed, rewarmed and evenly mixed. Serum and plasma samples can be frozen and thawed twice

at most.

7. Sample volume (**for Getein1100/Getein1160/Getein1180**): **10 µL**.

TEST PROCEDURE

1. Collect specimens according to user manual.
2. Test card, sample and reagent should reach to room temperature before test.

For Getein1100:

1. Confirm SD card lot No. in accordance with test kit lot No. Perform "SD card" calibration when necessary.
2. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.
3. Put the test card on a clean table, horizontally placed.
4. Using sample transfer pipette, deliver **10 µL** of sample into one tube of sample diluent, mix gently and thoroughly. Then drop **100µL** of sample mixture into the sample port on the test card.
5. **Reaction time: 10 minutes.** Insert the test card into Getein1100 and click on "Start" icon (for Android Getein1100) after reaction time is elapsed. The result will be shown on the screen and printed automatically.

For Getein1160/Getein1180:

1. Confirm SD card lot No. in accordance with test kit lot No.. Perform "SD card" calibration when necessary.
2. Enter testing interface of Getein1160/Getein1180.
3. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.
4. Put the test card on a clean table, horizontally placed.
5. Using sample transfer pipette, deliver **10 µL** of sample into one tube of sample diluent, mix gently and thoroughly. Then drop **100µL** of sample mixture into the sample port on the test card.
6. **Reaction time: 10 minutes.** Insert the test card into Getein1160/Getein1180 immediately after sample loading. The analyzer will count down the reaction time and automatically test the card after reaction time is elapsed. The result will be shown on the screen and printed automatically.

For Getein1200/Getein1600:

1. Each cartridge for Getein1200/Getein1600 contains a specific RFID card which can calibrate automatically.
2. Put the sample diluent at the correct position of Getein1200/Getein1600.
3. Place samples in the designed area of the sample holder, insert the holder and select the right test item, Getein1200/Getein1600 will do the testing and print the result automatically.

Notes:

1. It is required to perform "SD card" calibration when using

- a new batch of kits.
- It is suggested to calibrate once for one batch of kits for Getein1100/Getein1160/Getein1180.
 - Make sure the test card and the sample insertion are correct and complete.

TEST RESULTS

Getein1100/Getein1160/Getein1180/Getein1200/Getein1600 can scan the test card automatically and display the result on the screen. For additional information, please refer to the user manual of Getein1100/Getein1160/Getein1180/Getein1200/Getein1600.

Others: Samples whose concentration exceeds the upper limit should be diluted no more than 4 times.

EXPECTED VALUE

The expected normal value for ASO is determined by testing samples from 282 apparently healthy individuals. The upper 80th percentile value is 214.0 IU/mL. The upper 95th percentile value is 408.0 IU/mL. It is recommended that each laboratory determine the applicability of the reference value through experiments and establish its own reference ranges if necessary.

PERFORMANCE CHARACTERISTICS

Measuring Range	60.0-1370.0 IU/mL
Limit of Blank	≤40.0 IU/mL
Within-run Precision	≤10%
Between-run Precision	≤15%

LIMITATIONS

- Bilirubin and triglyceride in the sample may interfere with the test results, and the maximum allowable concentrations are 0.1 mg/mL and 10 mg/mL respectively.
- The test results of this kit are for clinical reference only, and should not be used as the sole criteria for clinical diagnosis. It is recommended to conduct a comprehensive analysis on the condition in combination with symptoms/signs, history and other laboratory tests.

REFERENCES

- Hahn RG, Knox LM, Forman TA. Evaluation of poststreptococcal illness. [J]. American family physician, 2005, 71(10): 1949-54.
- Sahin MS, Yalcin MU, Kocyigit D. Prevalence of rheumatic heart disease in patients with recurrent tonsillitis and elevated anti-streptolysin O titers. [J]. International Journal of Pediatric Otorhinolaryngology, 2016, 89: 133-5.

- Chang ST, Ku CH, Cherng SC. Evidence-based correlation between anti-streptolysin O serum titer and sacroiliac joint disorder. [J]. The Journal of rheumatology, 2007, 34(8): 1746-52.
- Ota H, Sato A, Matsumoto H, et. al. Immunological analysis of pseudo-positive reaction in healthy adults at anti-streptolysin O measurement by latex agglutination method. [J]. Rinsho byori. The Japanese journal of clinical pathology, 2005, 53(4): 279-83.

DESCRIPTION OF SYMBOLS USED

The following graphical symbols used in or found on the test kit are the most common ones appearing on medical devices and their packaging. They are explained in more details in the European Standard EN ISO 15223-1:2021.

Key to symbols used			
	Manufacturer		Use-by date
	Do not re-use		Date of manufacture
	Consult instructions for use or consult electronic instructions for use		Batch code
	Temperature limit		In vitro diagnostic medical device
	Contains sufficient for <n> tests		Authorized representative in the European Community/ European Union
	CE mark		Do not use if package is damaged and consult instructions for use
	Catalogue number		

Thank you for purchasing ASO Fast Test Kit (Immunofluorescence Assay).

Please read this user manual carefully before operating to ensure proper use.

Version: WIF83-S-05



Getein Biotech, Inc.
 Add.: No.9 Bofu Road, Luhe District, Nanjing, 211505, China
 Tel: +86-25-68568508
 Fax: +86-25-68568500
 E-mail: tech@getein.com.cn
 overseas@getein.com.cn
 Website: www.getein.com



CMC Medical Devices & Drugs S.L.
 Add.: C/ Horacio Lengo Nº 18, CP 29006, Málaga, Spain
 Tel: +34951214054